
Midmark M11 - v1.0.0
Total Cycles: 5054

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2024 04:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.6 F 0.0
2:00 188.6 F 3.1
4:00 203.3 F 8.3
6:00 224.5 F 10.0
8:00 227.9 F 8.9
10:00 238.7 F 10.4
12:00 241.5 F 12.1
14:00 240.6 F 10.8
16:00 254.3 F 17.6
18:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.7
2:00 272.2 F 29.1
3:00 271.9 F 28.5
4:00 272.4 F 29.2

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 216.0 F 0.0
10:00 228.4 F 0.0
15:00 230.7 F 0.0
20:00 232.5 F 0.0
25:00 229.5 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

11/04/2024 05:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____