
Midmark M11 - v1.0.0
Total Cycles: 4445

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2024 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 77.9 F 0.2
4:00 112.9 F 0.9
6:00 166.2 F 2.8
8:00 174.4 F 5.6
10:00 197.0 F 8.8
12:00 207.0 F 12.8
14:00 212.6 F 3.6
16:00 223.0 F 6.7
18:00 230.6 F 10.2
20:00 236.6 F 9.8
22:00 235.6 F 9.6
24:00 241.7 F 10.2
26:00 244.0 F 12.4
28:00 255.2 F 18.2
30:00 265.2 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.2
2:00 271.8 F 28.4
3:00 272.1 F 29.1
4:00 272.4 F 28.9

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 219.4 F 0.0
10:00 223.5 F 0.0
15:00 225.7 F 0.0
20:00 232.6 F 0.0
25:00 222.0 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:58
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:12
CYCLE COMPLETE

04/11/2024 09:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____