
Midmark M11 - v1.0.0
Total Cycles: 2725

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2022 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 201.2 F 3.2
4:00 201.8 F 8.0
6:00 223.1 F 13.4
8:00 229.9 F 8.4
10:00 239.9 F 13.5
12:00 240.3 F 11.3
14:00 239.0 F 9.9
16:00 252.8 F 16.6
18:00 264.3 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 29.0
2:00 271.7 F 28.4
3:00 271.3 F 28.6
4:00 272.4 F 29.0

Min 271.0 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 228.9 F 0.0
10:00 230.1 F 0.0
15:00 228.9 F 0.0
20:00 229.7 F 0.0
25:00 227.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

12/28/2022 01:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____