
Midmark M11 - v1.0.0
Total Cycles: 1545

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 185.2 F 4.2
4:00 198.2 F 8.9
6:00 222.7 F 8.6
8:00 228.2 F 9.3
10:00 235.4 F 9.3
12:00 242.1 F 12.4
14:00 241.6 F 11.1
16:00 255.1 F 17.9
18:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.7
2:00 272.3 F 29.1
3:00 271.8 F 28.4
4:00 272.6 F 29.1

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 227.1 F 0.0
10:00 242.1 F 0.0
15:00 237.8 F 0.0
20:00 231.4 F 0.0
25:00 219.9 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

01/20/2022 12:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____