
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000815
15:54 12/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:50
136.4 C 220kPa 3:50
136.4 C 225kPa 4:20
136.4 C 221kPa 4:50
136.2 C 224kPa 5:20
136.4 C 226kPa 5:50
136.1 C 221kPa 6:20
136.4 C 223kPa 6:50
136.4 C 225kPa 7:20
136.4 C 225kPa 7:20
VENTING 7:20
AIR DRYING 8:14
CYCLE COMPLETE 8:20

Digital Signature #
ABE8ED24B765BB09
