

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4611

Name: RITTER 1  
Sterilizer ID: V2178198  
06/05/2024 02:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.9 F 0.0  
2:00 116.2 F 1.2  
4:00 132.0 F 3.7  
6:00 164.0 F 6.8  
8:00 187.4 F 10.2  
10:00 201.2 F 6.0  
12:00 213.0 F 5.9  
14:00 224.8 F 9.8  
16:00 232.2 F 9.8  
18:00 236.6 F 10.4  
20:00 236.0 F 8.4  
22:00 246.7 F 13.8  
24:00 257.8 F 19.8  
26:00 268.1 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.9 F 29.2  
2:00 271.7 F 28.5  
3:00 273.0 F 29.4  
4:00 272.1 F 28.7

Min 271.6 F 28.5  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 201.9 F 0.0  
10:00 196.8 F 0.0  
15:00 198.3 F 0.0  
20:00 205.4 F 0.0  
25:00 205.0 F 0.0  
30:00 205.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 27:21  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:51  
CYCLE COMPLETE

06/05/2024 03:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_