
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001432
13:51 09/14/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 12:17
STERILIZING 13:50
136.0 C 220kPa 13:50
136.3 C 221kPa 14:20
136.4 C 223kPa 14:50
136.2 C 220kPa 15:20
136.3 C 222kPa 15:50
136.4 C 223kPa 16:21
136.1 C 221kPa 16:51
136.4 C 218kPa 17:21
136.4 C 223kPa 17:51
136.5 C 223kPa 18:21
136.3 C 222kPa 18:51
136.2 C 221kPa 19:21
136.4 C 223kPa 19:51
VENTING 19:51
AIR DRYING 20:49
CYCLE COMPLETE 50:49

Digital Signature #
CE871A4187857CDA
