
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002564
10:09 07/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:32
STERILIZING 3:29
136.0 C 220kPa 3:29
136.4 C 225kPa 3:59
136.3 C 224kPa 4:29
136.3 C 224kPa 4:59
136.4 C 223kPa 5:29
136.4 C 224kPa 5:59
136.2 C 223kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 6:59
VENTING 6:59
AIR DRYING 7:51
CYCLE COMPLETE 33:04

Digital Signature #
01BA64E94136A1C3
