
Midmark M11 - v1.0.0
Total Cycles: 2199

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2022 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 182.7 F 3.4
4:00 186.2 F 7.8
6:00 203.7 F 12.6
8:00 218.2 F 7.0
10:00 231.5 F 12.2
12:00 234.2 F 9.0
14:00 239.1 F 9.8
16:00 245.8 F 13.3
18:00 257.6 F 19.7
20:00 268.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.3
2:00 272.3 F 28.9
3:00 271.4 F 28.4
4:00 272.1 F 29.2

Min 271.3 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 210.4 F 0.0
10:00 235.1 F 0.0
15:00 226.2 F 0.0
20:00 228.0 F 0.0
25:00 230.2 F 0.0
30:00 231.8 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

08/04/2022 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____