
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000151
10:28 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:52
STERILIZING 3:14
136.4 C 220kPa 3:14
136.3 C 220kPa 3:44
136.2 C 220kPa 4:14
136.3 C 223kPa 4:44
136.4 C 224kPa 5:14
136.3 C 222kPa 5:44
136.2 C 222kPa 6:14
136.4 C 224kPa 6:44
136.4 C 224kPa 6:44
VENTING 6:44
AIR DRYING 7:41
CYCLE COMPLETE 7:44

Digital Signature #
F708C501D1E51EAC
