
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002787
14:43 06/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:47
136.4 C 224kPa 4:47
136.2 C 223kPa 5:17
136.3 C 224kPa 5:47
136.4 C 225kPa 6:17
136.2 C 219kPa 6:47
136.3 C 223kPa 7:17
136.1 C 220kPa 7:47
136.4 C 224kPa 8:17
136.4 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:16
CYCLE COMPLETE 9:18

Digital Signature #
EC6555E4E4C9F1B9
