
Midmark M11 - v1.0.5
Total Cycles: 1178

Name: RITTER2
Sterilizer ID: V2313228
06/09/2022 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 76.0 F 0.0
4:00 117.1 F 0.0
6:00 160.7 F 0.0
8:00 184.2 F 0.4
10:00 199.5 F 3.5
12:00 214.9 F 7.2
14:00 226.1 F 10.9
16:00 229.5 F 6.5
18:00 236.3 F 9.9
20:00 238.6 F 9.7
22:00 243.1 F 12.4
24:00 240.6 F 11.1
26:00 253.4 F 17.4
28:00 264.2 F 23.9
30:00 271.1 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 272.8 F 29.6
2:00 272.7 F 29.8
3:00 271.6 F 29.4
4:00 272.4 F 29.3

Min 271.5 F 29.1
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 238.4 F 0.0
10:00 219.6 F 0.0
15:00 227.9 F 0.0
20:00 222.7 F 0.0
25:00 222.4 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:37
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

06/09/2022 07:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____