
Midmark M11 - v1.0.5
Total Cycles: 4445

Name: RITTER2
Sterilizer ID: V2313228
03/24/2025 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 72.4 F 0.1
4:00 106.7 F 0.5
6:00 124.4 F 2.6
8:00 142.4 F 5.7
10:00 169.1 F 9.3
12:00 186.1 F 13.1
14:00 204.4 F 4.2
16:00 216.3 F 7.5
18:00 231.2 F 11.3
20:00 229.4 F 6.7
22:00 239.3 F 12.0
24:00 237.6 F 10.0
26:00 249.8 F 15.9
28:00 260.6 F 22.0
30:00 270.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.7 F 29.7
2:00 272.1 F 29.7
3:00 272.4 F 29.5
4:00 272.3 F 29.9

Min 271.6 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 191.8 F 0.0
10:00 214.3 F 0.0
15:00 214.8 F 0.0
20:00 213.1 F 0.0
25:00 214.9 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:09:25
CYCLE COMPLETE

03/24/2025 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____