

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1500

Name: RITTER 1  
Sterilizer ID: V2178198  
01/06/2022 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 75.8 F 0.0  
4:00 93.5 F 0.0  
6:00 121.0 F 0.0  
8:00 154.8 F 0.5  
10:00 193.4 F 3.5  
12:00 202.1 F 6.8  
14:00 215.0 F 10.4  
16:00 224.5 F 8.5  
18:00 231.4 F 8.6  
20:00 241.6 F 13.0  
22:00 239.5 F 10.3  
24:00 236.0 F 8.1  
26:00 250.2 F 15.2  
28:00 261.9 F 21.9  
30:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.9 F 28.4  
2:00 272.5 F 28.8  
3:00 272.7 F 28.8  
4:00 272.7 F 28.7

Min 271.6 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.5 F 0.0  
Dry Cycle Aborted

FILLING: 0:60  
HEATING: 30:16  
STERILIZING: 4:00  
VENTING: 2:47  
DRYING: 1:15  
TOTAL CYCLE: 00:39:17

01/06/2022 09:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_