

-----  
Midmark M11 - v1.0.5  
Total Cycles: 134

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2021 02:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.0 F 0.0  
2:00 164.9 F 3.7  
4:00 184.7 F 8.9  
6:00 211.8 F 6.8  
8:00 223.5 F 10.8  
10:00 235.4 F 9.0  
12:00 236.1 F 9.4  
14:00 249.4 F 15.5  
16:00 260.1 F 22.9  
18:00 272.6 F 30.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 30.4  
1:00 273.0 F 30.5  
2:00 272.1 F 30.5  
3:00 271.7 F 30.3  
4:00 272.0 F 30.5  
  
Min 271.6 F 30.3  
Max 273.4 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 240.2 F 0.0  
10:00 217.4 F 0.0  
15:00 219.6 F 0.0  
20:00 229.5 F 0.0  
25:00 208.5 F 0.0  
30:00 200.6 F 0.0  
  
Drying Complete

FILLING: 2:13  
HEATING: 18:13  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 00:57:24  
CYCLE COMPLETE

06/23/2021 03:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_