
Midmark M11 - v1.0.5
Total Cycles: 4136

Name: RITTER2
Sterilizer ID: V2313228
11/19/2024 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.5 F 0.0
2:00 146.3 F 0.1
4:00 169.0 F 2.6
6:00 188.7 F 6.4
8:00 214.9 F 10.4
10:00 220.3 F 6.2
12:00 230.0 F 9.7
14:00 234.5 F 9.5
16:00 241.0 F 12.3
18:00 241.0 F 11.2
20:00 253.1 F 17.5
22:00 264.0 F 24.0
24:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 30.0
2:00 271.8 F 29.7
3:00 272.0 F 29.4
4:00 272.5 F 29.8

Min 271.0 F 28.9
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 218.5 F 0.0
10:00 202.1 F 0.0
15:00 198.2 F 0.0
20:00 194.7 F 0.0
25:00 198.7 F 0.0
30:00 202.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:38
CYCLE COMPLETE

11/19/2024 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____