
Midmark M11 - v1.0.5
Total Cycles: 2708

Name: RITTER2
Sterilizer ID: V2313228
07/27/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.7 F 0.0
2:00 176.0 F 1.2
4:00 177.5 F 4.8
6:00 204.2 F 9.8
8:00 226.5 F 6.8
10:00 232.2 F 10.5
12:00 230.9 F 7.6
14:00 242.8 F 12.4
16:00 244.4 F 12.7
18:00 256.8 F 19.3
20:00 267.6 F 26.0
22:00 273.0 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 273.2 F 29.8
2:00 272.2 F 29.8
3:00 272.5 F 29.7
4:00 272.2 F 29.9

Min 271.7 F 29.3
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 232.8 F 0.0
10:00 221.4 F 0.0
15:00 227.1 F 0.0
20:00 226.5 F 0.0
25:00 227.2 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

07/27/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____