

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001048  
10:10 29/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:53  
PRESSURIZING 2:58  
STERILIZING 4:04  
136.0 C 220kPa 4:04  
136.2 C 223kPa 4:34  
136.3 C 223kPa 5:04  
136.3 C 223kPa 5:34  
136.2 C 224kPa 6:04  
136.1 C 223kPa 6:34  
136.4 C 225kPa 7:04  
136.4 C 224kPa 7:34  
136.4 C 224kPa 7:34  
VENTING 7:34  
AIR DRYING 8:19  
CYCLE COMPLETE 23:19  
-----

Digital Signature #  
A635266978172A89  
-----