

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4463

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2025 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 83.4 F 0.0  
4:00 75.4 F 0.5  
6:00 98.8 F 2.6  
8:00 134.4 F 4.9  
10:00 162.7 F 7.6  
12:00 181.6 F 10.7  
14:00 211.1 F 5.7  
16:00 217.0 F 6.2  
18:00 226.1 F 9.7  
20:00 234.8 F 11.8  
22:00 235.0 F 9.8  
24:00 238.0 F 9.9  
26:00 244.6 F 13.0  
28:00 255.0 F 18.7  
30:00 264.4 F 24.6  
32:00 271.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.5  
1:00 272.1 F 29.9  
2:00 271.2 F 29.9  
3:00 273.9 F 30.7  
4:00 272.3 F 30.7  
  
Min 270.9 F 29.4  
Max 273.9 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
5:00 165.0 F 0.0  
10:00 173.5 F 0.0  
15:00 173.1 F 0.0  
20:00 171.6 F 0.0  
25:00 169.5 F 0.0  
30:00 176.4 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:05  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:10:42  
CYCLE COMPLETE

03/27/2025 10:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_