

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006475  
12:37 16/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:31  
STERILIZING 3:39  
136.4 C 223kPa 3:39  
136.4 C 225kPa 4:09  
136.3 C 223kPa 4:39  
136.0 C 222kPa 5:09  
136.2 C 223kPa 5:39  
136.4 C 222kPa 6:09  
136.4 C 222kPa 6:39  
136.3 C 223kPa 7:09  
136.3 C 223kPa 7:09  
VENTING 7:09  
AIR DRYING 7:59  
CYCLE COMPLETE 31:46

-----  
Digital Signature #  
A0A76D73CEA0B321  
-----