

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001746  
8:40 14/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:21  
PRESSURIZING 1:26  
STERILIZING 2:14  
136.0 C 220kPa 2:14  
136.2 C 220kPa 2:44  
136.3 C 222kPa 3:14  
136.0 C 219kPa 3:44  
136.4 C 224kPa 4:14  
136.4 C 221kPa 4:44  
136.4 C 223kPa 5:14  
136.3 C 221kPa 5:44  
136.3 C 221kPa 5:44  
VENTING 5:44  
AIR DRYING 6:35  
CYCLE COMPLETE 6:36

-----  
Digital Signature #  
CDF32885D55CB62D  
-----