
Midmark M11 - v1.0.0
Total Cycles: 3986

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2023 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.0 F 0.0
2:00 163.6 F 2.7
4:00 175.4 F 6.2
6:00 198.3 F 10.6
8:00 220.2 F 4.8
10:00 223.1 F 8.0
12:00 234.5 F 12.3
14:00 235.6 F 8.7
16:00 243.1 F 12.0
18:00 244.0 F 12.2
20:00 255.6 F 18.1
22:00 265.8 F 24.3
24:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.1 F 28.6
2:00 272.8 F 28.9
3:00 271.7 F 28.8
4:00 272.1 F 28.6

Min 271.4 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 219.9 F 0.0
10:00 199.7 F 0.0
15:00 187.6 F 0.0
20:00 184.3 F 0.0
25:00 184.3 F 0.0
30:00 186.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

11/06/2023 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____