
Midmark M11 - v1.0.5
Total Cycles: 1194

Name: RITTER2
Sterilizer ID: V2313228
06/16/2022 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 169.9 F 2.5
4:00 175.7 F 6.4
6:00 202.4 F 11.6
8:00 222.7 F 6.4
10:00 236.1 F 11.7
12:00 233.6 F 8.9
14:00 239.8 F 10.5
16:00 245.4 F 13.6
18:00 258.0 F 20.1
20:00 268.4 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.7 F 29.5
2:00 271.6 F 29.5
3:00 272.7 F 29.7
4:00 272.0 F 29.9

Min 271.0 F 29.0
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 233.6 F 0.0
10:00 222.3 F 0.0
15:00 226.5 F 0.0
20:00 224.6 F 0.0
25:00 155.9 F 0.0
30:00 152.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

06/16/2022 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____