
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001895
8:40 26/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:22
STERILIZING 3:24
136.0 C 220kPa 3:24
135.9 C 222kPa 3:54
136.2 C 223kPa 4:24
136.4 C 224kPa 4:54
136.4 C 225kPa 5:24
136.2 C 223kPa 5:54
136.4 C 225kPa 6:24
136.4 C 225kPa 6:54
136.4 C 225kPa 6:54
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 37:45

Digital Signature #
C2AA7F832643A4EC
