
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001646
15:36 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:41
136.4 C 223kPa 3:41
136.1 C 223kPa 4:11
136.0 C 222kPa 4:41
136.1 C 222kPa 5:11
136.2 C 223kPa 5:41
136.4 C 224kPa 6:11
136.4 C 225kPa 6:41
136.4 C 225kPa 7:11
136.4 C 225kPa 7:11
VENTING 7:11
AIR DRYING 7:59
CYCLE COMPLETE 8:03

Digital Signature #
E28D145E1E93C108
