

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4844

Name: RITTER 1  
Sterilizer ID: V2178198  
08/22/2024 02:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.4 F 0.0  
2:00 122.6 F 0.3  
4:00 133.9 F 0.9  
6:00 179.6 F 1.7  
8:00 198.4 F 5.4  
10:00 212.7 F 9.2  
12:00 229.7 F 13.2  
14:00 228.6 F 7.4  
16:00 238.7 F 11.9  
18:00 235.5 F 8.7  
20:00 241.8 F 10.9  
22:00 245.4 F 12.9  
24:00 256.8 F 19.0  
26:00 267.3 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.3 F 29.1  
2:00 272.6 F 29.0  
3:00 271.9 F 28.6  
4:00 271.3 F 28.4

Min 271.3 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 187.1 F 0.0  
10:00 184.9 F 0.0  
15:00 184.3 F 0.0  
20:00 187.6 F 0.0  
25:00 208.8 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 27:39  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:11  
CYCLE COMPLETE

08/22/2024 03:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_