
Midmark M11 - v1.0.5
Total Cycles: 551

Name: RITTER2
Sterilizer ID: V2313228
11/15/2021 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.9 F 0.0
2:00 189.7 F 4.6
4:00 201.5 F 10.7
6:00 216.0 F 6.1
8:00 232.9 F 12.2
10:00 236.1 F 10.5
12:00 236.9 F 9.8
14:00 251.8 F 17.2
16:00 264.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.4 F 29.7
2:00 272.7 F 30.2
3:00 271.8 F 29.4
4:00 272.6 F 29.9

Min 271.8 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 202.0 F 0.0
10:00 205.4 F 0.0
15:00 205.5 F 0.0
20:00 202.6 F 0.0
25:00 201.6 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 17:51
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:30
CYCLE COMPLETE

11/15/2021 01:31 PM
ACCEPTED: X
REJECTED:
Operator: U02