

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4699

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2024 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.5 F 0.0  
2:00 142.2 F 1.5  
4:00 144.9 F 4.0  
6:00 166.9 F 6.9  
8:00 198.8 F 10.4  
10:00 216.3 F 6.9  
12:00 222.5 F 6.7  
14:00 226.3 F 10.7  
16:00 230.9 F 7.4  
18:00 237.4 F 11.1  
20:00 236.4 F 8.5  
22:00 248.0 F 14.5  
24:00 258.8 F 20.4  
26:00 268.3 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.5 F 29.4  
2:00 271.7 F 28.7  
3:00 271.4 F 28.9  
4:00 272.4 F 29.3

Min 271.1 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 211.6 F 0.0  
10:00 211.5 F 0.0  
15:00 208.4 F 0.0  
20:00 211.2 F 0.0  
25:00 209.3 F 0.0  
30:00 208.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 27:20  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:06:08  
CYCLE COMPLETE

07/05/2024 10:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_