
Midmark M11 - v1.0.5
Total Cycles: 2149

Name: RITTER2
Sterilizer ID: V2313228
03/15/2023 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 150.4 F 1.8
4:00 158.4 F 4.7
6:00 183.5 F 9.0
8:00 202.0 F 8.0
10:00 221.9 F 8.2
12:00 233.3 F 13.5
14:00 237.6 F 10.7
16:00 235.3 F 8.9
18:00 249.3 F 15.7
20:00 261.0 F 22.4
22:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.4
1:00 271.8 F 29.7
2:00 272.8 F 30.2
3:00 272.4 F 29.7
4:00 271.6 F 29.3

Min 271.5 F 29.2
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 209.8 F 0.0
10:00 216.4 F 0.0
15:00 208.5 F 0.0
20:00 204.9 F 0.0
25:00 206.6 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

03/15/2023 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____