

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3325

Name: RITTER 1  
Sterilizer ID: V2178198  
06/02/2023 07:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	74.4 F	0.0
2:00	78.0 F	0.0
4:00	108.8 F	0.0
6:00	128.8 F	0.0
8:00	157.5 F	0.0
10:00	180.0 F	0.1
12:00	194.3 F	3.1
14:00	206.1 F	5.8
16:00	215.4 F	8.7
18:00	225.3 F	12.0
20:00	226.0 F	6.1
22:00	236.6 F	10.3
24:00	237.0 F	9.1
26:00	244.7 F	12.6
28:00	242.2 F	11.1
30:00	254.0 F	17.0
32:00	264.4 F	23.2
34:00	271.6 F	27.8

mm:ss	Degrees	PSI
0:00	271.4 F	27.9
1:00	272.7 F	28.6
2:00	271.7 F	27.9
3:00	272.1 F	28.5
4:00	272.4 F	28.4

Min 271.4 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	174.1 F	0.0
5:00	201.3 F	0.0
10:00	208.8 F	0.0
15:00	209.7 F	0.0
20:00	208.7 F	0.0
25:00	208.1 F	0.0
30:00	206.7 F	0.0

Drying Complete

FILLING: 0:45  
HEATING: 34:07  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:12:20  
CYCLE COMPLETE

06/02/2023 09:05 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_