

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3143

Name: RITTER 1  
Sterilizer ID: V2178198  
04/21/2023 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.7 F 0.0  
2:00 150.7 F 0.0  
4:00 179.7 F 0.0  
6:00 201.9 F 2.2  
8:00 217.2 F 6.9  
10:00 229.0 F 11.3  
12:00 230.5 F 7.2  
14:00 242.5 F 13.0  
16:00 241.3 F 11.2  
18:00 239.7 F 10.1  
20:00 254.0 F 17.2  
22:00 266.2 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.5 F 28.9  
2:00 272.0 F 28.2  
3:00 272.2 F 28.7  
4:00 272.5 F 28.6

Min 271.5 F 28.0  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 213.8 F 0.0  
10:00 231.5 F 0.0  
15:00 228.4 F 0.0  
20:00 229.7 F 0.0  
25:00 229.1 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:01:32  
CYCLE COMPLETE

04/21/2023 11:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_