

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001430  
11:16 09/14/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:14  
PRESSURIZING 7:24  
STERILIZING 9:10  
136.0 C 220kPa 9:10  
136.5 C 221kPa 9:41  
136.3 C 220kPa 10:11  
136.0 C 218kPa 10:41  
136.5 C 219kPa 11:11  
136.5 C 222kPa 11:41  
136.2 C 220kPa 12:11  
136.3 C 221kPa 12:40  
VENTING 12:40  
AIR DRYING 13:48  
CYCLE COMPLETE 43:48  
-----

Digital Signature #  
5478C969346FAAB8  
-----