

-----  
Midmark M11 - v1.0.0  
Total Cycles: 876

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2021 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.4 F 0.0  
2:00 115.0 F 0.0  
4:00 129.7 F 0.1  
6:00 164.9 F 0.2  
8:00 187.1 F 2.8  
10:00 201.4 F 5.8  
12:00 214.2 F 9.3  
14:00 226.9 F 13.3  
16:00 230.2 F 8.1  
18:00 240.8 F 12.7  
20:00 238.4 F 10.0  
22:00 239.2 F 9.4  
24:00 248.4 F 14.1  
26:00 259.4 F 20.1  
28:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.6  
2:00 271.5 F 27.9  
3:00 272.7 F 28.7  
4:00 272.2 F 28.2

Min 271.4 F 27.8  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 209.8 F 0.0  
10:00 220.2 F 0.0  
15:00 227.4 F 0.0  
20:00 224.5 F 0.0  
25:00 224.9 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:05  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:07:20  
CYCLE COMPLETE

06/09/2021 10:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_