
Midmark M11 - v1.0.5
Total Cycles: 1983

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 79.3 F 0.0
4:00 99.7 F 0.0
6:00 128.1 F 0.0
8:00 161.9 F 0.1
10:00 182.3 F 3.1
12:00 203.7 F 6.2
14:00 217.5 F 10.0
16:00 226.7 F 9.3
18:00 230.4 F 9.4
20:00 236.8 F 10.5
22:00 240.4 F 11.7
24:00 239.4 F 10.2
26:00 251.9 F 16.5
28:00 263.0 F 23.0
30:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.8 F 29.5
2:00 272.8 F 30.0
3:00 272.0 F 29.2
4:00 272.8 F 29.7

Min 271.9 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 212.0 F 0.0
10:00 217.3 F 0.0
15:00 231.1 F 0.0
20:00 224.9 F 0.0
25:00 227.3 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:58
CYCLE COMPLETE

01/31/2023 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____