
Midmark M11 - v1.0.0
Total Cycles: 1904

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2022 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 184.1 F 3.9
4:00 199.1 F 9.0
6:00 218.0 F 7.1
8:00 229.8 F 9.9
10:00 234.2 F 8.1
12:00 244.7 F 13.6
14:00 244.7 F 12.6
16:00 258.1 F 19.8
18:00 269.9 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.2
4:00 272.2 F 28.5

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 231.8 F 0.0
10:00 230.0 F 0.0
15:00 236.3 F 0.0
20:00 222.4 F 0.0
25:00 220.3 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

05/03/2022 04:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____