
Midmark M11 - v1.0.0
Total Cycles: 4404

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.8 F	0.0
2:00	71.4 F	0.2
4:00	75.2 F	1.0
6:00	91.4 F	2.7
8:00	122.3 F	5.0
10:00	150.4 F	7.6
12:00	170.9 F	10.3
14:00	191.1 F	13.2
16:00	203.7 F	3.4
18:00	213.6 F	6.4
20:00	223.3 F	9.5
22:00	233.0 F	13.2
24:00	233.9 F	9.1
26:00	243.3 F	12.5
28:00	242.0 F	11.5
30:00	252.7 F	16.8
32:00	262.0 F	22.3
34:00	271.0 F	28.2

STERILIZING

mm:ss	Degrees	PSI
0:00	271.1 F	28.6
1:00	272.3 F	29.4
2:00	272.0 F	28.8
3:00	271.0 F	28.6
4:00	272.3 F	29.5

Min 271.0 F 28.4
Max 272.4 F 29.5

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
0:00	199.6 F	0.0	
5:00	230.7 F	0.0	
10:00	239.3 F	0.0	
15:00	229.3 F	0.0	
20:00	228.1 F	0.0	
25:00	228.6 F	0.0	
30:00	230.7 F	0.0	

Drying Complete

FILLING: 0:36
HEATING: 34:51
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:13:10
CYCLE COMPLETE

03/26/2024 09:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____