

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000759  
14:06 27/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:48  
STERILIZING 3:57  
136.0 C 227kPa 3:57  
136.1 C 231kPa 4:27  
136.4 C 231kPa 4:57  
136.3 C 232kPa 5:27  
136.4 C 233kPa 5:57  
136.2 C 231kPa 6:27  
136.3 C 231kPa 6:57  
136.4 C 233kPa 7:27  
136.4 C 233kPa 7:27  
VENTING 7:27  
AIR DRYING 8:22  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
8D693A0F498C8550  
-----