
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001469
9:47 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:26
136.1 C 221kPa 4:26
136.4 C 225kPa 4:56
136.4 C 225kPa 5:26
136.2 C 222kPa 5:56
136.3 C 224kPa 6:26
136.2 C 224kPa 6:56
136.4 C 225kPa 7:26
136.2 C 222kPa 7:56
136.2 C 222kPa 7:56
VENTING 7:56
AIR DRYING 8:50
CYCLE COMPLETE 9:05

Digital Signature #
332E7BBCFA782826
