
Midmark M11 - v1.0.0
Total Cycles: 3228

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 07:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 75.8 F 0.0
4:00 102.2 F 0.0
6:00 131.7 F 0.0
8:00 157.2 F 0.1
10:00 186.5 F 3.2
12:00 210.3 F 6.6
14:00 207.6 F 10.2
16:00 221.2 F 9.0
18:00 229.1 F 7.4
20:00 239.0 F 11.5
22:00 233.2 F 7.7
24:00 245.7 F 13.3
26:00 243.0 F 11.5
28:00 254.7 F 17.6
30:00 265.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.5 F 28.9
2:00 271.5 F 28.0
3:00 272.6 F 28.8
4:00 271.4 F 28.1

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 230.1 F 0.0
10:00 225.5 F 0.0
15:00 227.8 F 0.0
20:00 234.1 F 0.0
25:00 230.9 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:46
CYCLE COMPLETE

05/15/2023 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____