
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007092
8:07 25/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:47
STERILIZING 3:47
136.4 C 223kPa 3:47
136.2 C 223kPa 4:17
136.2 C 222kPa 4:47
136.2 C 221kPa 5:17
136.4 C 223kPa 5:47
136.4 C 223kPa 6:17
136.2 C 221kPa 6:47
136.3 C 221kPa 7:17
136.3 C 221kPa 7:17
VENTING 7:17
AIR DRYING 8:09
CYCLE COMPLETE 8:22

Digital Signature #
2EB6E8E6D9C51AC6
