
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000007
7:33 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:11
PRESSURIZING 1:17
STERILIZING 2:20
136.4 C 223kPa 2:20
135.9 C 222kPa 2:50
136.3 C 224kPa 3:20
136.1 C 220kPa 3:50
136.4 C 223kPa 4:20
136.4 C 223kPa 4:50
136.4 C 221kPa 5:20
136.4 C 220kPa 5:50
136.4 C 220kPa 5:50
VENTING 5:50
AIR DRYING 6:40
CYCLE COMPLETE 6:48

Digital Signature #
D345D62BEECD1923
