
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000892
14:28 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 3:59
STERILIZING 5:12
136.4 C 221kPa 5:12
136.3 C 231kPa 5:42
136.3 C 232kPa 6:12
136.4 C 233kPa 6:42
136.4 C 231kPa 7:12
136.4 C 231kPa 7:42
136.4 C 232kPa 8:12
136.4 C 233kPa 8:42
136.4 C 233kPa 8:42
VENTING 8:42
AIR DRYING 9:36
CYCLE COMPLETE 16:31

Digital Signature #
213281DB3C8CFC61
