
Midmark M11 - v1.0.0
Total Cycles: 3525

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2023 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.1 F 0.0
2:00 189.1 F 3.7
4:00 206.8 F 8.3
6:00 222.1 F 9.7
8:00 230.9 F 9.3
10:00 238.6 F 9.7
12:00 243.4 F 12.8
14:00 242.7 F 11.8
16:00 256.3 F 18.9
18:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 271.6 F 28.7
3:00 272.5 F 29.0
4:00 271.6 F 28.4

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 229.9 F 0.0
10:00 228.8 F 0.0
15:00 222.0 F 0.0
20:00 223.1 F 0.0
25:00 223.8 F 0.0
30:00 223.7 F 0.0

Drying Complete

FILLING: 1:54
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

07/13/2023 01:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____