
Midmark M11 - v1.0.5
Total Cycles: 1824

Name: RITTER2
Sterilizer ID: V2313228
12/08/2022 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.5 F 0.0
2:00 199.7 F 3.8
4:00 215.4 F 9.3
6:00 223.8 F 7.8
8:00 234.2 F 11.1
10:00 235.0 F 9.0
12:00 238.6 F 9.6
14:00 249.4 F 15.2
16:00 261.6 F 22.3
18:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 271.5 F 29.1
4:00 272.6 F 29.4

Min 271.1 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 235.1 F 0.0
10:00 226.1 F 0.0
15:00 224.2 F 0.0
20:00 222.1 F 0.0
25:00 218.9 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

12/08/2022 11:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____