

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002643  
15:21 31/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:06  
PRESSURIZING 2:15  
STERILIZING 3:24  
136.5 C 226kPa 3:24  
136.4 C 226kPa 3:54  
136.4 C 225kPa 4:24  
136.2 C 222kPa 4:54  
136.4 C 225kPa 5:24  
136.4 C 226kPa 5:54  
136.4 C 223kPa 6:24  
136.2 C 222kPa 6:54  
136.2 C 222kPa 6:54  
VENTING 6:54  
AIR DRYING 7:46  
CYCLE COMPLETE 11:55

-----  
Digital Signature #  
59E2E88F1AE9FB31  
-----