
Midmark M11 - v1.0.0
Total Cycles: 2789

Name: RITTER 1
Sterilizer ID: V2178198
01/18/2023 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.6 F 0.0
2:00 80.2 F 0.0
4:00 95.6 F 0.0
6:00 121.3 F 0.0
8:00 157.3 F 1.4
10:00 181.1 F 3.9
12:00 197.5 F 6.9
14:00 210.8 F 10.2
16:00 223.7 F 9.2
18:00 227.2 F 7.8
20:00 237.3 F 12.2
22:00 236.0 F 9.1
24:00 239.7 F 10.0
26:00 246.4 F 13.4
28:00 258.0 F 19.7
30:00 268.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.7
2:00 272.4 F 29.1
3:00 271.8 F 28.3
4:00 272.7 F 29.1

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 220.0 F 0.0
10:00 226.2 F 0.0
15:00 222.1 F 0.0
20:00 227.6 F 0.0
25:00 228.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:06
CYCLE COMPLETE

01/18/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____