

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000246  
13:29 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:33  
STERILIZING 5:14  
136.3 C 224kPa 5:14  
136.4 C 233kPa 5:44  
136.3 C 233kPa 6:14  
136.1 C 231kPa 6:44  
136.3 C 232kPa 7:14  
136.3 C 232kPa 7:44  
136.4 C 230kPa 8:14  
136.4 C 233kPa 8:44  
136.4 C 233kPa 8:44  
VENTING 8:44  
AIR DRYING 9:46  
CYCLE COMPLETE 12:41  
-----

Digital Signature #  
D3F31750623A1BCA  
-----