
Midmark M11 - v1.0.5
Total Cycles: 877

Name: RITTER2
Sterilizer ID: V2313228
03/14/2022 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 165.4 F 3.2
4:00 184.1 F 7.4
6:00 206.0 F 12.7
8:00 221.0 F 7.6
10:00 234.6 F 13.1
12:00 237.5 F 10.8
14:00 237.1 F 9.6
16:00 251.1 F 16.3
18:00 263.2 F 23.4
20:00 271.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.6
2:00 271.3 F 29.0
3:00 272.6 F 29.8
4:00 272.0 F 29.2

Min 271.3 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 219.2 F 0.0
10:00 209.9 F 0.0
15:00 206.2 F 0.0
20:00 203.9 F 0.0
25:00 234.0 F 0.0
30:00 232.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

03/14/2022 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____