
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004547
14:54 17/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:32
STERILIZING 2:32
136.0 C 220kPa 2:32
136.4 C 224kPa 3:02
136.2 C 222kPa 3:32
136.1 C 220kPa 4:02
136.3 C 224kPa 4:32
136.4 C 224kPa 5:02
136.4 C 225kPa 5:32
136.4 C 224kPa 6:02
136.4 C 224kPa 6:02
VENTING 6:02
AIR DRYING 7:19
CYCLE COMPLETE 7:37

Digital Signature #
98916281F31468B3
