
Midmark M11 - v1.0.5
Total Cycles: 1249

Name: RITTER2
Sterilizer ID: V2313228
07/04/2022 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 172.1 F 3.1
4:00 186.1 F 7.3
6:00 209.8 F 12.4
8:00 221.3 F 7.0
10:00 233.2 F 12.0
12:00 234.8 F 9.4
14:00 240.0 F 9.7
16:00 247.0 F 14.3
18:00 258.7 F 20.8
20:00 269.4 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.8
2:00 271.7 F 29.0
3:00 272.2 F 29.9
4:00 272.4 F 29.5

Min 271.4 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 215.8 F 0.0
10:00 226.0 F 0.0
15:00 232.8 F 0.0
20:00 225.8 F 0.0
25:00 222.7 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

07/04/2022 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____